

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00545	Calib. Date:	10/16/2015
Certification File No.:	00546	Cert. Date:	10/16/2015
Linearity File No.:	00547	Lin. Date:	10/16/2015
Solution File No.:	00569	Soln. Date:	02/07/2016
Sequential File No.:	00569	File Date:	02/07/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0217
Solution Control Lot:	15H140	Expires:	08/04/2017
		Bottle No.:	1350

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:21S		
Control 1 EC	0.097%	00:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	00:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:23S		
Control 2 EC	0.097%	00:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	00:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:24S		
Control 3 EC	0.096%	00:25S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	00:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:26S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JOYCE

First Name: PATRICK

MI: D

Signature: _____



Badge No.: 326

Date: 02/07/2016