

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0056
Location:	LONG BRANCH POLICE DEPT	
Calibration File No.:	00746	Calib. Date: 12/13/2018
Certification File No.:	00747	Calib. No.: 00029
Linearity File No.:	00748	Cert. Date: 12/13/2018
Solution File No.:	00766	Cert. No.: 00023
Sequential File No.:	00766	Lin. Date: 12/13/2018
		Lin. No.: 00023
		Soln. Date: 03/11/2019
		Soln. No.: 00157
		File Date: 03/11/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXN S3-0216
Solution Control Lot:	17200	Expires: 06/28/2019
		Bottle No.: 1413

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:06S		
Control 1 EC	0.099%	08:07S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:07S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:08S		
Control 2 EC	0.100%	08:09S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	08:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:09S		
Control 3 EC	0.098%	08:10S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:11S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: VALDES

First Name: HEATHER

MI: -

Signature: _____

Badge No.: 290

Date: 03/11/2019