

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0056
Location:	LONG BRANCH POLICE DEPT	
Calibration File No.:	00629	Calib. Date: 10/21/2016
Certification File No.:	00630	Calib. No.: 00025
Linearity File No.:	00631	Cert. Date: 10/21/2016
Solution File No.:	00661	Cert. No.: 00019
Sequential File No.:	00661	Lin. Date: 10/21/2016
		Lin. No.: 00019
		Soln. Date: 03/17/2017
		Soln. No.: 00135
		File Date: 03/17/2017

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDXN S3-0217
Control Solution %:	0.100%		Expires: 08/04/2017
Solution Control Lot:	15H140		Bottle No.: 1336

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:51S		
Control 1 EC	0.101%	09:52S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:53S		
Control 2 EC	0.101%	09:54S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:54S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:55S		
Control 3 EC	0.101%	09:55S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:56S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HOLLIS-VALDES First Name: HEATHER

MI: -

Signature: _____

Badge No.: 290

Date: 03/17/2017