

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00593	Calib. Date:	04/22/2016
Certification File No.:	00594	Cert. No.:	00024
Linearity File No.:	00595	Cert. Date:	04/22/2016
Solution File No.:	00596	Lin. Date:	04/22/2016
Sequential File No.:	00596	Lin. No.:	00018
		Soln. Date:	04/22/2016
		Soln. No.:	00120
		File Date:	04/22/2016

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXN S3-0216
Control Solution %:	0.100%			Expires:	08/04/2017
Solution Control Lot:	15H140			Bottle No.:	1338

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:40D		
Control 1 EC	0.100%	10:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:42D		
Control 2 EC	0.100%	10:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:43D		
Control 3 EC	0.100%	10:44D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:44D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXAP2-114 (DK)

Changed By:

Last Name: KLIMIK

First Name: DAVID

MI: W

Signature: TPR D. KLIMIK #7040

Badge No.: 7040

Date: 04/22/2016