

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00746	Calib. Date:	12/13/2018
Certification File No.:	00747	Calib. No.:	00029
Linearity File No.:	00748	Cert. Date:	12/13/2018
Solution File No.:	00771	Cert. No.:	00023
Sequential File No.:	00771	Lin. Date:	12/13/2018
		Lin. No.:	00023
		Soln. Date:	05/04/2019
		Soln. No.:	00159
		File Date:	05/04/2019

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXN S3-0216
Control Solution %:	0.100%			Expires:	06/28/2019
Solution Control Lot:	17200			Bottle No.:	0456

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:58D		
Control 1 EC	0.099%	05:59D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	05:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:00D		
Control 2 EC	0.098%	06:00D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:01D		
Control 3 EC	0.098%	06:02D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OTERO

First Name: JESUS

MI: C

Signature: _____



Badge No.: 340

Date: 05/04/2019