

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00665 Calib. Date: 04/04/2017 Calib. No.: 00026
Certification File No.: 00666 Cert. Date: 04/04/2017 Cert. No.: 00020
Linearity File No.: 00667 Lin. Date: 04/04/2017 Lin. No.: 00020
Solution File No.: 00676 Soln. Date: 05/11/2017 Soln. No.: 00138
Sequential File No.: 00676 File Date: 05/11/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0217
Control Solution %: 0.100% Expires: 08/04/2017
Solution Control Lot: 15H140 Bottle No.: 1331

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:11D		
Control 1 EC	0.103%	02:12D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:13D		
Control 2 EC	0.101%	02:14D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	02:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:15D		
Control 3 EC	0.099%	02:15D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	02:15D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SAMOL

First Name: GEORGE

MI: L

Signature: _____



Badge No.: 330

Date: 05/11/2017