

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00713 Calib. Date: 07/17/2018 Calib. No.: 00028
Certification File No.: 00714 Cert. Date: 07/17/2018 Cert. No.: 00022
Linearity File No.: 00715 Lin. Date: 07/17/2018 Lin. No.: 00022
Solution File No.: 00726 Soln. Date: 09/26/2018 Soln. No.: 00150
Sequential File No.: 00726 File Date: 09/26/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0217
Control Solution %: 0.100% Expires: 06/28/2019
Solution Control Lot: 17200 Bottle No.: 0293

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:41D		
Control 1 EC	0.101%	12:42D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	12:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:43D		
Control 2 EC	0.101%	12:43D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	12:43D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:44D		
Control 3 EC	0.101%	12:45D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CHAPARRO

First Name: SERGIO

MI: A

Signature: _____

Badge No.: 352

Date: 09/26/2018