

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0056  
Location: LONG BRANCH POLICE DEPT  
Calibration File No.: 00746 Calib. Date: 12/13/2018 Calib. No.: 00029  
Certification File No.: 00747 Cert. Date: 12/13/2018 Cert. No.: 00023  
Linearity File No.: 00748 Lin. Date: 12/13/2018 Lin. No.: 00023  
Solution File No.: 00754 Soln. Date: 01/11/2019 Soln. No.: 00155  
Sequential File No.: 00754 File Date: 01/11/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0216  
Control Solution %: 0.100% Expires: 06/28/2019  
Solution Control Lot: 17200 Bottle No.: 1342

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:26S		
Control 1 EC	0.098%	05:27S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	05:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:28S		
Control 2 EC	0.098%	05:29S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	05:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:30S		
Control 3 EC	0.098%	05:31S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	05:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: OLSKI

First Name: GEORGE

MI: T

Signature: \_\_\_\_\_

Badge No.: 336

Date: 01/11/2019