

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT	Calib. No.:	00028
Calibration File No.:	00713	Cert. Date:	07/17/2018
Certification File No.:	00714	Lin. Date:	07/17/2018
Linearity File No.:	00715	Soln. Date:	10/10/2018
Solution File No.:	00731	File Date:	10/10/2018
Sequential File No.:	00731		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXN S3-0217
Control Solution %:	0.100%			Expires:	06/28/2019
Solution Control Lot:	17200			Bottle No.:	1344

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:11D		
Control 1 EC	0.099%	10:12D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:12D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:13D		
Control 2 EC	0.099%	10:13D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:14D		
Control 3 EC	0.100%	10:15D	33.8°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:15D	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WALLS

First Name: CHRISTOPHER

MI:

Signature: _____



Badge No.: 314

Date: 10/10/2018