

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0056  
Location: LONG BRANCH POLICE DEPT  
Calibration File No.: 00713 Calib. Date: 07/17/2018 Calib. No.: 00028  
Certification File No.: 00714 Cert. Date: 07/17/2018 Cert. No.: 00022  
Linearity File No.: 00715 Lin. Date: 07/17/2018 Lin. No.: 00022  
Solution File No.: 00738 Soln. Date: 11/05/2018 Soln. No.: 00152  
Sequential File No.: 00738 File Date: 11/05/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0217  
Control Solution %: 0.100% Expires: 06/28/2019  
Solution Control Lot: 17200 Bottle No.: 1405

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:33S		
Control 1 EC	0.101%	16:33S	33.8°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:33S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:34S		
Control 2 EC	0.101%	16:35S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	16:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:36S		
Control 3 EC	0.100%	16:37S	33.8°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:37S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:38S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


### Changed By:

Last Name: OLSKI

First Name: GEORGE

MI: T

Signature: \_\_\_\_\_



Badge No.: 336

Date: 11/05/2018