

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00629	Calib. Date:	10/21/2016
Certification File No.:	00630	Cert. No.:	00019
Linearity File No.:	00631	Lin. Date:	10/21/2016
Solution File No.:	00637	Soln. No.:	00128
Sequential File No.:	00637	File Date:	11/09/2016
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0217
Solution Control Lot:	15H140	Expires:	08/04/2017
		Bottle No.:	1344

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:50S		
Control 1 EC	0.102%	03:50S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	03:50S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:51S		
Control 2 EC	0.101%	03:52S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:53S		
Control 3 EC	0.101%	03:53S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	03:53S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:54S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: OLSKI

First Name: GEORGE

MI: T

Signature: \_\_\_\_\_



Badge No.: 336

Date: 11/09/2016