Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT				
Calibration File No.:	00682	Calib. Date:	09/14/2017	Calib. No.:	00027
Certification File No.:	00683	Cert. Date:	09/14/2017	Cert. No.:	00021
Linearity File No.:	00684	Lin. Date:	09/14/2017	Lin. No.:	00021
Solution File No.:	00698	Soln. Date:	11/16/2017	Soln. No.:	00143
Sequential File No.:	00698	File Date:	11/16/2017		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXN S3-0216
Control Solution %:	0.100%			Expires:	06/28/2019
Solution Control Lot:	17200			Bottle No.:	1382

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:10S		
Control 1 EC	0.100%	08:11S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:11S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:12S		
Control 2 EC	0.100%	08:13S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:13S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:14S		
Control 3 EC	0.100%	08:14S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:14S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXA P2-114 (DK)

Changed By:

Last Name: KLIMIK First Name: DAVID MI: W

Badge No.: 7040

Signature: Tpr II D. Klimik #7040 Date: 11/16/2017