

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARXA-0056
Location:	LONG BRANCH POLICE DEPT	
Calibration File No.:	00713	Calib. Date: 07/17/2018
Certification File No.:	00714	Cert. No.: 00028
Linearity File No.:	00715	Cert. Date: 07/17/2018
Solution File No.:	00745	Lin. Date: 07/17/2018
Sequential File No.:	00745	Lin. No.: 00022
		Soln. Date: 12/09/2018
		Soln. No.: 00153
		File Date: 12/09/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXN S3-0217
Solution Control Lot:	17200	Expires: 06/28/2019
		Bottle No.: 1343

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:55S		
Control 1 EC	0.101%	08:56S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	08:56S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:57S		
Control 2 EC	0.101%	08:57S	33.8°C	*** TEST PASSED ***
Control 2 IR	0.102%	08:57S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:58S		
Control 3 EC	0.101%	08:59S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:00S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WALLS

First Name: CHRIS

MI:

Signature: _____

Badge No.: 314

Date: 12/09/2018