

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00629	Calib. Date:	10/21/2016
Certification File No.:	00630	Cert. No.:	00025
Linearity File No.:	00631	Cert. Date:	10/21/2016
Solution File No.:	00646	Lin. Date:	10/21/2016
Sequential File No.:	00646	Lin. No.:	00019
		Soln. Date:	12/13/2016
		Soln. No.:	00130
		File Date:	12/13/2016

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXN S3-0217
Control Solution %:	0.100%			Expires:	08/04/2017
Solution Control Lot:	15H140			Bottle No.:	1327

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:07S		
Control 1 EC	0.101%	04:07S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:07S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:08S		
Control 2 EC	0.100%	04:09S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:10S		
Control 3 EC	0.100%	04:11S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	04:11S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:12S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OLSKI

First Name: GEORGE

MI: T

Signature: _____



Badge No.: 336

Date: 12/13/2016