Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT				
Calibration File No.:	00746	Calib. Date:	12/13/2018	Calib. No.:	00029
Certification File No.:	00747	Cert. Date:	12/13/2018	Cert. No.:	00023
Linearity File No.:	00748	Lin. Date:	12/13/2018	Lin. No.:	00023
Solution File No.:	00749	Soln. Date:	12/13/2018	Soln. No.:	00154
Sequential File No.:	00749	File Date:	12/13/2018		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXN S3-0216
Control Solution %:	0.100%			Expires:	06/28/2019
Solution Control Lot:	17200			Bottle No.:	1345

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:48S		
Control 1 EC	0.099%	13:48S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:49S		
Control 2 EC	0.098%	13:50S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:51S		
Control 3 EC	0.099%	13:52S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXA P2-114 DN

Changed By:

Last Name: NAPOLITANO First Name: DAVID MI: M.

Badge No.: 7237

Signature: 7237 Date: 12/13/2018