

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0056
Location: LONG BRANCH POLICE DEPT
Calibration File No.: 00682 Calib. Date: 09/14/2017 Calib. No.: 00027
Certification File No.: 00683 Cert. Date: 09/14/2017 Cert. No.: 00021
Linearity File No.: 00684 Lin. Date: 09/14/2017 Lin. No.: 00021
Solution File No.: 00701 Soln. Date: 12/19/2017 Soln. No.: 00144
Sequential File No.: 00701 File Date: 12/19/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXN S3-0216
Control Solution %: 0.100% Expires: 06/28/2019
Solution Control Lot: 17200 Bottle No.: 1396

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:31S		
Control 1 EC	0.100%	21:32S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	21:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:33S		
Control 2 EC	0.101%	21:33S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:34S		
Control 3 EC	0.098%	21:35S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:36S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ESKRIDGE

First Name: STANLEY

MI: J

Signature: _____

 #313

Badge No.: 313

Date: 12/19/2017