

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00746	Calib. Date:	12/13/2018
Certification File No.:	00747	Calib. No.:	00029
Linearity File No.:	00748	Cert. Date:	12/13/2018
Solution File No.:	00761	Lin. Date:	12/13/2018
Sequential File No.:	00761	Lin. No.:	00023
		Soln. Date:	02/12/2019
		Soln. No.:	00156
		File Date:	02/12/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXN S3-0216
Solution Control Lot:	17200	Expires:	06/28/2019
		Bottle No.:	0632

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:23S		
Control 1 EC	0.099%	02:24S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:25S		
Control 2 EC	0.098%	02:26S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	02:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:27S		
Control 3 EC	0.096%	02:28S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:29S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: OLSKI

First Name: GEORGE

MI: T

Signature: _____



336

Badge No.: 336

Date: 02/12/2019