

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0056
Location:	LONG BRANCH POLICE DEPT		
Calibration File No.:	00682	Calib. Date:	09/14/2017
Certification File No.:	00683	Cert. Date:	09/14/2017
Linearity File No.:	00684	Lin. Date:	09/14/2017
Solution File No.:	00685	Soln. Date:	09/14/2017
Sequential File No.:	00685	File Date:	09/14/2017
		Calib. No.:	00027
		Cert. No.:	00021
		Lin. No.:	00021
		Soln. No.:	00140

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDXN S3-0217
Control Solution %:	0.100%			Expires:	06/28/2019
Solution Control Lot:	17200			Bottle No.:	1437

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:32D		
Control 1 EC	0.101%	11:33D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:34D		
Control 2 EC	0.100%	11:34D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:35D		
Control 3 EC	0.099%	11:36D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DEEPZ-090 (KA)

Changed By:

Last Name: ALCOTT	First Name: KEVIN	MI: W.
Signature: <u>Tpt. I K W. #6704</u>	Badge No.: 6704	Date: 09/14/2017